


Cention® & Cention® Primer

Indications requiring defect-oriented preparation (with Cention Primer)



Indications requiring retentive preparation



Retentive preparation guidelines



Manufacturer



Ivoclar Vivadent AG, 9494 Schaan/Liechtenstein

www.ivoclarvivadent.com

ivoclar
vivadent®
clinical

Cention® & Cention® Primer

Cention®
Primer



Cention®



Mixing time

Dispensing time 20s

Working time 2:00 min

Setting time 6:30 min

CE 0123

See
Instructions
for Use



ivoclar
vivadent®
clinical